

# Nuclear imaging and semi-invasive electrocardiography in CRT

**Prof. Dr. med. Andreas Otte,  
Facharzt für Nuklearmedizin**

Fakultät Elektrotechnik  
und Informationstechnik (E+I)

Badstraße 24  
77652 Offenburg  
Tel. 0781 205-338  
E-Mail: andreas.otte@hs-offenburg.de

**1987-1994:** Studium der Physik (1987-88), Universität Konstanz; Studium der Humanmedizin (1988-1994), Albert-Ludwigs-Universität Freiburg; Promotion zum Dr. med.

**1994-1999:** Weiterbildung zum Facharzt für Nuklearmedizin, Universitätskliniken Basel und Freiburg

**1999-2010:** Leitende Positionen in der klinischen Forschung (u.a. Bereichsleiter ZNS bei Gödecke/Pfizer, Freiburg; Stv. Leiter des Zentrums Klinische Studien, Universitätsklinikum Freiburg)

**Lehre:** 2002-2009: Visiting Professor Universität Gent/Belgien; seit 2007 Honorarprofessor Hochschule Lahr; seit 2010 Professor für biomedizinische Systemtechnik, Hochschule Offenburg

**Forschungsgebiet:** Neurosysteme; biomedizinische bildgebende Verfahren



## 1.4 Nuclear imaging and semi-invasive electrocardiography in CRT

*Prof. Dr. med. Andreas Otte  
Prof. Dr. rer. nat. habil. Bruno Ismer  
PD Dr. Ing. med. habil. Matthias Heinke  
Prof. Dr. med. Juraj Melicherik*

*Dr. Heinke is from the Department of Internal Medicine I, Friedrich-Schiller-University of Jena, and Professor Melicherik is from the Mediclin Heartcenter Lahr*

### Abstract

*Cardiac resynchronisation therapy (CRT) is a promising treatment option in patients with chronic heart failure. In this article the roles of semi-invasive esophageal left-heart electrocardiography and functional cardiac nuclear imaging in the field of CRT are highlighted, as the combination of both could be a favourable diagnostic approach in special cardiac situations. Also original esophageal left heart electrogram data of exemplary CRT patients is presented.*

### Introduction

During the last 15 years cardiac resynchronisation therapy (CRT) has made a huge step forward from experimental pacing to a routine and evidence-based treatment option in patients with chronic heart failure, who have ischemic or non-ischemic cardiomyopathy at NYHA III or IV, sinus rhythm, a left-bundle branch block and a left ventricular ejection fraction below 35% [1, 2].

The relatively young field of biomedical engineering has also found a rapid and promising development in the two medical imaging disciplines radiology and nuclear medicine as well as semi-inva-

sive electrocardiographic methods and thus strengthened modern medicine.

### The role of nuclear imaging

Nuclear medicine diagnostic instruments provide various opportunities to help the patient in the battle against his or her disease. In cardiovascular diseases, electrocardiogram (ECG)-triggered myocardial perfusion single-photon-emission-computed-tomography (SPECT) or <sup>18</sup>F-fluorodeoxy-D-glucose (FDG) positron emission tomography (PET) have become routine tools to detect viable myocardium, to study heart motility and to define the ejection fraction, the left-ventricular or right-ventricular function.

Although CRT is a promising treatment option, some patients do not respond to it [1]. van Campen et al., 2007, therefore, studied the relationship between viable myocardial tissue assessed by FDG PET and the improvement after CRT in 39 patients with ischemic cardiomyopathy, ejection fraction 27±9% and a wide QRS complex [3]. They found a correlation between the extent of viable myocardium and the hemodynamic response to CRT (assessed by the cardiac index measured by Doppler echocardiography) and concluded that FDG PET imaging may be a useful predicting tool to distinguish between responders and non-responders. Siegrist et al., 2008, studied the role of left-ventricular phase histograms of radionuclide ventriculography (RVN) for optimizing atrioventricular delay setting for CRT in 17 patients with heart failure and could show that RVN is a helpful instrument for the optimization of pacing parameters in CRT [4]. Looking at radiological tools, selection of pacing sites can be improved e.g.


by cardiac computer tomography of the cardiac veins. In addition to PET imaging, candidates with large myocardial scars can be excluded also by magnetic resonance tomography [5].

### The role of semi-invasive left-heart electrocardiography


Recently, esophageal left heart electrogram was demonstrated to be a simple but useful semi-invasive method in the field of CRT [6, 7]. It could be utilized to increase responder rate (Figure 1.4-1). Performed preoperatively, this approach provides more direct quantification of the left ventricular conduction delay than QRS width in surface ECG. Duncan et. al. found an underestimation of duration of ventricular activation by the 12-lead ECG compared with pacemaker device electrogram measurements [8]. Previously measured esophageal left ventricular conduction delay can be used, intraoperatively, as valuable target interval to guide positioning of the left ventricular electrode. Missing this target interval should encourage the seeking for another pacing position in order to decrease the amount of nonresponders (Figure 1.4-2).

Furthermore, postoperatively, esophageal left atrial electrogram recording enables measurement of implant-related interatrial conduction intervals. The latter mainly determine individual duration of the hemodynamically optimal AV delay programming.

Actually, the semi-invasive esophageal left heart electrogram is a feature of the Biotronik ICS3000 (Biotronik GmbH, Berlin, Germany) standard programmer, solely. Applying, perorally, a TOSlim esophageal electrode (Osypka AG,



**Fig. 1.4-1:** Example of CRT patient presenting distinct esophageal left ventricular conduction delay Q-ELV of 112 ms. In contrast to that, the left ventricular electrode is placed in unfavourable region with left ventricular electrode delay Q-LV of 60 ms, only. Intraoperatively, this suboptimal situation should encourage the seeking for another left ventricular electrode position in order to avoid nonresponders



**Fig. 1.4-2:** Example of CRT patient demonstrating the esophageal left ventricular conduction delay Q-ELV of 54 ms as minimal target interval for left ventricular electrode positioning. In this constellation with Q-LV of 74 ms, response to CRT could be expected

Rheinfelden-Herten, Germany) connected to the PK199 programmer cable, depending on esophageal electrode position, filtered esophageal left ventricular or left atrial electrograms can be recorded simultaneously with three-channel surface ECG in patients without or with pacing systems irrespective of their make and model.

### Conclusion

Nuclear imaging as well as semi-invasive left-heart electrogram studies in CRT are still rare, and the number of patients studied is low. Both methods, however, can help CRT in special situations. Therefore, further clinical studies combining the two disciplines may be a favourable approach.

### References

[1] Abraham WT, Fisher WG, Smith AL et al.: MIRACLE Study Group: Multi-center InSync Randomized Clinical Evaluation. Cardiac resynchronization in chronic heart failure. 346: 1845 – 53, N Engl J Med 2002

[2] Götze S, Butter C, Fleck E.: Kardiale Resynchronisation bei Herzinsuffizienz [Cardiac resynchronization therapy for heart failure]. 95 (suppl. 4): 1 – 18, Clin Res Cardiol 2006

[3] van Campen CM, Visser FC, van der Weerd AP et al.: FDG PET as a predictor of response to resynchronization therapy in patients with ischaemic cardiomyopathy. 34: 309 – 15 Eur J Nucl Med Mol Imaging 2007

[4] Siegrist PT, Comte N, Holzmeister J et al.: Effects of AV delay programming on ventricular resynchronization: role of radionuclide ventriculography. 35: 1516 – 22, Eur J Nucl Med Mol Imaging 2008

[5] Flachskampf FA. The non-responders in cardiac resynchronization therapy: do we have solution? European Society of Cardiology, Congress Report 2011: link: <http://www.escardio.org/congresses/esc-2011/congress-reports/Pages/101-Flachskampf.aspx> [retrieved: Nov 22, 2011]

[6] Ismer B, Kleimenhagen F, Heinke M et al.: Significance of new electrocardiographic parameters to improve cardiac resynchronization therapy. 56(Suppl. 1). Walter de Gruyter, Berlin, Boston: DOI 10.1515/BMT.2011.512, Biomed Tech 2011

[7] Heinke M, Ismer B, Kuehnert H et al.: Transesophageal left ventricular electrogram-recording and temporary pacing to improve patient selection for cardiac resynchronization. Med Biol Eng Comput 2011; 49: 851 – 8. DOI 10.1007/s11517-011-0767-6

[8] Duncan AM, Lim E, Mebrate Y et al.: Underestimation of duration of ventricular activation by 12-lead ECG compared with direct measurements of activation duration derived from implanted pacemaker leads. 152: 35 – 42, Intern J Cardiol 2011